

ITS.

Implants for Trauma Surgery



Hand
Locking Plates
System

THE ART of TRAUMA SURGERY

The Art of Trauma Surgery is a collaborative project between I.T.S. and Austrian artist Oskar Stocker that celebrates the skill, perseverance, and artistry of surgeons and engineers who work tirelessly to improve outcomes for trauma patients.

At I.T.S., we stand for long-term, trusting relationships with our customers, suppliers, and development partners. Through our devotion to innovation and development, we continuously seek to improve and optimize products and techniques in the field of traumatology.

We believe that the success of our mission lies in the combination of the technical expertise, compassion and dedication of surgeons and engineers to help patients regain their health and well-being. Join us in celebrating these remarkable individuals and *The Art of Trauma Surgery!*

About the Artist

The Austrian artist Oskar Stocker (b. 1956) lives and works in Graz, Austria. He has become known internationally through the exhibition *Facing Nations*, which consists of portraits of more than 120 people of various nationalities living in Graz; it was shown first in Graz itself, then in Vienna, and later culminated in 2010 with its display at the UN Headquarters in New York City.

In addition to the portraits of individual people, he devotes himself to the depiction of landscapes and objects, down to the smallest detail.



All I.T.S. plates are preformed anatomically as a matter of principle. If adjustment of the plate to the shape of the bone is required, this is possible by carefully bending gently in one direction once. Particular care is required when bending in the region of a plate hole, as deformation of the plate may lead to a failure of the locking mechanism. The plate must not be buckled or bent several times. This is particularly important in the case of titanium implants, to prevent material fatigue and subsequent failure. The method of bending is the conscious responsibility of the operating doctor; I.T.S. GmbH can accept no liability whatsoever for this.

Table of Contents

I. Introduction

- 8 Plate Technology
- 9 System Overview
- 10 Properties
- 12 Indications
- 12 Contraindications
- 12 Time of Operation

Introduction

2. Surgical Technique

- 16 Pre-operative Patient Preparation
- 16 Access
- 16 Reduction
- 17 **OPTIONAL:** Temporary Plate Fixation
- 17 Screw Placement
- 19 Postoperative Treatment
- 19 Explantation

Surgical Technique

3. Information

- 22 Technical Information
- 23 Type II Anodization
- 24 Ordering Information

Information



Introduction




○ Plate Technology

At I.T.S., we stand for long-term, trusting relationships with our customers, suppliers, and development partners. Through our dedication to innovation and development, we continuously seek to improve and optimize products and techniques for trauma surgery.

ONE Technology for all implants

All I.T.S. plates are made from Titanium Grade 2, whereas the screws are made of a harder titanium alloy. This allows the plates to have only non-threaded holes, which all (with the exception of oblong holes) accept both non-locking and locking screws.




When a locking screw is inserted, it forms threads into the plate. There is no cutting and thus no debris is created. Each locking screw can be locked at a free placement within a cone of angulation up to $\pm 15^\circ$ and can be re-positioned up to three times.


○ System Overview

The ITS. Hand Locking Plates System is a trusted osteosynthesis solution, offering a variety of plate types to address a range of hand fractures.

A special feature of this system is the freely selectable hole configuration, which makes it possible to choose locking or non-locking screws for each hole (except compression and oblong holes). With an angle selection of $+\text{- } 15^\circ$, the system offers particular versatility, especially for complex fractures.



Properties



● Screws

37151-xx

LOCKING

Stabilization Screw, D=1.5mm
Spiral Drill, D=1.1mm
Torque, T5



37251-xx

LOCKING

Stabilization Screw, D=2.3mm
Spiral Drill, D=1.8mm
Torque, T5



32151-xx

NON-LOCKING

Cortical Screw, D=1.8mm
Spiral Drill, D=1.5mm
Torque, T5



● Indications

The ITS. HLS - Hand Locking Plates System is indicated for use in fracture fixation of:

- the phalanges
- the metacarpal bones
- the carpal bones
- for arthrodesis
- for corrective osteotomies and
- for subcapital radial head fractures

● Contraindications

- Existing infections in the fracture zone and operation area
- Common situations that do not allow osteosynthesis
- With advanced osteoporosis
- Skin and soft-tissue problems which prevent a tension-free closure of the skin
- Obesity
- Lack of patient compliance


● Time of Operation

- Immediately after trauma or delayed
- After regression of swelling

Intended purpose

The hand plate system - P32 is used to treat arthrodeses in the hand area as well as subcapital radial head fractures. In addition, it can also be used in the Shape of corrective osteotomies. Other indications include fractures of the phalanges, carpal and metacarpal bones.





Surgical Technique

2 ■

● Pre-operative Patient Preparation

- Elevated upper body approx. 30° - 40° inclination, shoulder freely movable (optional shoulder-table).
- The arm should be freely movable to have the possibility of fracture reduction
- General anaesthesia, regional anaesthesia or a combination can be used
- Possible use of medication for blood arrest

The following surgical steps are illustrated using the HLS Plate 2 T-Shape 1.0mm on the OS metacarpale V. and also apply to other plates.

For details on screw, drill, and screwdriver allocation, see *the screw overview p. II.*

● Access

Dorsal access to the middle finger joint: Straight skin incision centrally over the joint, careful dissection through the subcutaneous tissue. The extensor tendon is held aside paratendinously and the joint is then exposed.

IMPORTANT: The incision depends on the affected area of the hand!



● Reduction

- Temporary fixation of the fracture parts using forceps
- Subsequent control under fluoroscopy

● OPTIONAL: Temporary Plate Fixation

Optionally, the plate can be stabilized using the Temporary Plate Holder (58100-100).



● Screw Placement

- Use the spiral drill to drill through the drill guide, $D=1.2/1.9\text{mm}$ (62211) (drill diameter depends on the choice of screw - see page II).



- The screw length is determined using the depth gauge (9-110).



- Depending on the measured length, either locking stabilization screws or non-locking stable cortical screws (37151-XX or 37251-XX / 32151-XX) are used. The screws are inserted using the torque shank T5x80, AO Connector (54101-80), in combination with the AO handle (53013).



- Subsequent control of the plate and screw position under fluoroscopy.



○ Postoperative Treatment

- Elevation and preventative edema measures on the day of the operation
- Free weightbearing according to symptoms and stipulations of the operating surgeons

○ Explantation

Removal is possible, if desired by the patient. This is facilitated by the fact that, due to different materials of plate and screws, cold welding never occurs.

The ITS. Type II anodization surface treatment reduces the risk of cold welding of titanium implants (*for more information, see p. 23*).




work

Information

3.

○ Technical Information



For detailed cleaning and sterilization instructions, please refer to package insert.

Not true to scale

○ Type II Anodization


Chemical process - anodization in a strong alkaline solution*

Type III anodization

- Layer thickness 60-200nm
 - + Different colors
 - Implant surface remains sensitive to:
Chipping
Peeling
Discoloration

Type II anodization

- Layer thickness 1000-2000nm
 - + Film becomes an interstitial part of the titanium
 - No visible cosmetic effect



Anodization Type II leads to following benefits*

- Oxygen and silicon absorbing conversion layer
- Decrease in protein adsorption
- Closing of micro pores and micro cracks
- Reduced risk of inflammation and allergy
- Hardened titanium surface
- Reduced tendency of cold welding of titanium implants
- Increased fatigue resistance of implants
- Improved wear and friction characteristics

* White Paper: Ti6Al4V with Anodization Type II: Biological Behavior and Biomechanical Effects; Axel Baumann, Nils Zander

● Ordering Information

HLS Plates - Straight



21502

Description	Article Number
HLS Plate, Straight, Individual	21502

HLS Plates - Straight

1.0mm



21503-4



21503-5



21503-6



21503-8



21503-10

Description	Holes	Article Number
HLS Plates, Straight 1.0mm	4	21503-4
HLS Plates, Straight 1.0mm	5	21503-5
HLS Plates, Straight 1.0mm	6	21503-6
HLS Plates, Straight 1.0mm	8	21503-8
HLS Plates, Straight 1.0mm	10	21503-10

HLS Plates - Straight

1.0mm



2I504-4



2I504-5



2I504-6



2I504-8



2I504-10

Description	Holes	Article Number
HLS Plates, Straight 1.5mm	4	2I504-4
HLS Plates, Straight 1.5mm	5	2I504-5
HLS Plates, Straight 1.5mm	6	2I504-6
HLS Plates, Straight 1.5mm	8	2I504-8
HLS Plates, Straight 1.5mm	10	2I504-10

(Optional)

1.0mm



2I503-16

1.5mm




2I504-16

Description	Holes	Article Number
HLS Plates, Straight 1.0mm	16	2I503-16
HLS Plates, Straight 1.5mm	16	2I504-16


HLS Plates - T-Shape

1.0mm



21505-2 21505-3 21505-4

1.5mm



21506-2 21506-3 21506-4

Description	Holes	Article Number
HLS Plates, 2 T-Shape 1.0mm	2	21505-2
HLS Plates, 2 T-Shape 1.0mm	3	21505-3
HLS Plates, 2 T-Shape 1.0mm	4	21505-4
HLS Plates, 2 T-Shape 1.5mm	2	21506-2
HLS Plates, 2 T-Shape 1.5mm	3	21506-3
HLS Plates, 2 T-Shape 1.5mm	4	21506-4

(Optional)

1.0mm



21505-7

1.5mm



21506-8

Description	Holes	Article Number
HLS Plates, 2 T-Shape 1.0mm	7	21505-7
HLS Plates, 2 T-Shape 1.5mm	8	21506-8

1.0mm



2I507-2




2I507-3



2I507-4

1.5mm



2I508-2



2I508-3



2I508-4

Description	Holes	Article Number
HLS Plates, 3 T-Shape 1.0mm	2	2I507-2
HLS Plates, 3 T-Shape 1.0mm	3	2I507-3
HLS Plates, 3 T-Shape 1.0mm	4	2I507-4
HLS Plates, 3 T-Shape 1.5mm	2	2I508-2
HLS Plates, 3 T-Shape 1.5mm	3	2I508-3
HLS Plates, 3 T-Shape 1.5mm	4	2I508-4

(Optional)

1.0mm



2I507-7

1.5mm



2I508-8

Description	Holes	Article Number
HLS Plates, 3 T-Shape 1.0mm	7	2I507-7
HLS Plates, 3 T-Shape 1.5mm	8	2I508-8

HLS Plates - T-Shape

1.0mm



2I509-2



2I509-3



2I509-4

1.5mm



2I510-2



2I510-3



2I510-4

Description	Holes	Article Number
HLS Plates, 4 T-Shape 1.0mm	2	2I509-2
HLS Plates, 4 T-Shape 1.0mm	3	2I509-3
HLS Plates, 4 T-Shape 1.0mm	4	2I509-4
HLS Plates, 4 T-Shape 1.5mm	2	2I510-2
HLS Plates, 4 T-Shape 1.5mm	3	2I510-3
HLS Plates, 4 T-Shape 1.5mm	4	2I510-4

HLS Plates - T-Shape Extended




2I501

Description	Article Number
HLS Plates, T-Shape Extended	2I501

HLS Plates - Y-Shape

1.0mm




2I5I3-3

2I5I3-4

2I5I3-5

1.5mm



2I5I4-3

2I5I4-4

2I5I4-5

Description	Holes	Article Number
HLS Plates, Y-Shape 1.0mm	3	2I5I3-3
HLS Plates, Y-Shape 1.0mm	4	2I5I3-4
HLS Plates, Y-Shape 1.0mm	5	2I5I3-5
HLS Plates, Y-Shape 1.5mm	3	2I5I4-3
HLS Plates, Y-Shape 1.5mm	4	2I5I4-4
HLS Plates, Y-Shape 1.5mm	5	2I5I4-5

(Optional)

1.0mm



2I5I3-7

1.5mm



2I5I4-7


Description	Holes	Article Number
HLS Plates, Y-Shape 1.0mm	7	2I5I3-7
HLS Plates, Y-Shape 1.5mm	7	2I5I4-7

HLS Plates - L-Shape



Description	Article Number
HLS Plates, L-Shape	Left
HLS Plates, L-Shape	Right

(Optional)



Description	Holes	Article Number
HLS Plates, L-Shape	7	2I5I6-7
HLS Plates, L-Shape	7	2I5I5-7

HLS Plates - Square



2I5I7

Description	Article Number
HLS Plates, Square 1.0mm	2I5I7



Description	Article Number
HLS Plates, Square 1.5mm	Small
HLS Plates, Square 1.5mm	Large
HLS Plates, Square 1.5mm	X-Large

Screws

Stabilization Screw, D=1.5mm	Length	Article Number
Locking	5	3715I-5
	6	3715I-6
	7	3715I-7
	8	3715I-8
	9	3715I-9
	10	3715I-10
	11	3715I-11
	12	3715I-12
	13	3715I-13
	14	3715I-14
	15	3715I-15
	16	3715I-16
	18	3715I-18
	20	3715I-20
	22	3715I-22
	24	3715I-24
	26	3715I-26

Stabilization Screw, D=2.3mm	Length	Article Number
Locking	5	3725I-5
	6	3725I-6
	7	3725I-7
	8	3725I-8
	9	3725I-9
	10	3725I-10
	11	3725I-11
	12	3725I-12
	13	3725I-13
	14	3725I-14
	15	3725I-15
	16	3725I-16
	18	3725I-18
	20	3725I-20
	22	3725I-22
	24	3725I-24
	26	3725I-26

Cortical Screw, D=1.8mm	Length	Article Number
Non-Locking	5	3215I-5
	6	3215I-6
	7	3215I-7
	8	3215I-8
	9	3215I-9
	10	3215I-10
	11	3215I-11
	12	3215I-12
	13	3215I-13
	14	3215I-14
	15	3215I-15
	16	3215I-16
	18	3215I-18
	20	3215I-20

Instruments

(Optional) Plate Holder



58I00-I00

Description	Article Number
Plate Holder for Hand System	58I00-I00

Spiral Drill



6III3-60



9-0I2



6II83-I00

Description	Article Number
Spiral Drill, D=1.1mm, L=60mm, AO Connector	6III3-60
Spiral Drill, D=2.5mm, L=85mm, AO Connector	9-0I2
Spiral Drill, D=1.8mm, L=100mm, AO Connector	6II83-I00

Drill Guide



622II

Description	Article Number
Drill Guide, D=1.2/1.9mm	622II

Depth Gauge



9-II0

Description	Article Number
Depth Gauge	9-II0

AO-Silicone Handle



53013

Description	Article Number
AO-Silicone Handle	53013

Torque-Shank



54101-80



T5



54101-80-2

Description	Article Number
Torque-Shank, T5x80, AO Connector	54101-80
Self Holding Sleeve, Torque, T5 Shank	54101-80-2

Tweezer



HB 2001

Description	Article Number
Tweezer, Straight	HB 2001

Plate Holding Forceps



06-586

Description	Article Number
Plate Holding Forceps, 15.5cm	06-586

Bending Forceps



9-406

Description	Article Number
Bending Forceps	9-406


Pointed Forceps




9-596

Description	Article Number
Pointed Forceps	9-596

Forceps




9-646



62820

Description	Article Number
Forceps, Plate Holder	9-646
Forceps, Drill Guide	62820

Plate Cutter



KH.267.I6-L

Description	Article Number
Plate Cutter, Long	KH.267.I6-L

Notes

Notes

Disclaimer:

The intended users are limited to medical personnel with appropriate product training by the medical product consultants or knowledge of the surgical procedure to be applied. The medical staff must ensure that the use of I.T.S. GmbH medical devices is appropriate, taking into account the medical condition and medical history of the patient. Prior to product use, medical personnel must refer to complete information on product label and in IFU, including, but not limited to, indications, contraindications, warnings and preventative measures, and cleaning and sterilization instructions. Product availability is dependent on country registrations and clearances. For more information, please visit www.its-implant.com or contact us at office@its-implant.com. Unless otherwise noted, all information herein is the intellectual property of I.T.S. GmbH.



HEADQUARTER

I.T.S. GmbH

Autal 28, 8301 Lassnitzhöhe, Austria

Tel.: +43 (0) 316/ 211 21 0

office@its-implant.com

www.its-implant.com

I.T.S. USA

1778 Park Avenue N - Suite 200

Tel.: 407-971-8054

info@its-implantusa.com

www.its-implant.com

Order No. HLS-OP-I225-US

Edition: December/2025

© I.T.S. GmbH Graz/Austria 2025

Subject to technical alterations, errors and misprints excepted.